

ISLAND SCIENTIFIC LTD



*Vacuum Equipment, Used and Reconditioned for Technology,
Industry, Research and Education*

Reply to:

Unit 9
Ventnor Industrial Estate
Station Road
Ventnor
Isle of Wight
PO38 1DX
England

Tel No: (0) 1983 855822

Fax No: (0) 1983 852146

E-Mail Address: sales@island-scientific.co.uk

Web Site: <http://www.island-scientific.co.uk>



Make: Edwards

Type: Spectron 5000 E-NL

Detectable Leak:-

Massive: 9×10^{-2} - 1×10^{-4} mbar ls-1

Gross: 9×10^{-0} - 3×10^{-9} mbar ls-1

Fine: 9×10^{-3} - 2×10^{-1} mbar ls-1

Pumping System:-

Rotary Roughing: E2M18 fitted with internal EMF20 oil mist filter and standard roughing

Rotary Pump: E2M1.5 fitted with EMF10 oil mist filter and Visco probe connection port

Diffusion Pump: E203 & EO50, high sensitivity accumulator fitted

Dimensions:-

Length: 685 mm

Width: 580 mm

Height: 1200 mm